

RESEARCH

Open Access

Towards identification of finger flexions using single channel surface electromyography – able bodied and amputee subjects

Dinesh Kant Kumar^{1†}, Sridhar Poosapadi Arjunan^{1*†} and Vijay Pal Singh^{2†}

Abstract

Background: This research has established a method for using single channel surface electromyogram (sEMG) recorded from the forearm to identify individual finger flexion. The technique uses the volume conduction properties of the tissues and uses the magnitude and density of the singularities in the signal as a measure of strength of the muscle activity.

Methods: sEMG was recorded from the *flexor digitorum superficialis* muscle during four different finger flexions. Based on the volume conduction properties of the tissues, sEMG was decomposed into wavelet maxima and grouped into four groups based on their magnitude. The mean magnitude and the density of each group were the inputs to the twin support vector machines (TSVM). The algorithm was tested on 11 able-bodied and one trans-radial amputated volunteer to determine the accuracy, sensitivity and specificity. The system was also tested to determine inter-experimental variations and variations due to difference in the electrode location.

Results: Accuracy and sensitivity of identification of finger actions from single channel sEMG signal was 93% and 94% for able-bodied and 81% and 84% for trans-radial amputated respectively, and there was only a small inter-experimental variation.

Conclusions: Volume conduction properties based sEMG analysis provides a suitable basis for identifying finger flexions from single channel sEMG. The reported system requires supervised training and automatic classification.

Background

Surface electromyogram (sEMG) is the non-invasive recording of the electrical activity of the muscle. It is closely related to muscle contraction and an indicator of the associated actions. For an amputee, sEMG of the residual muscles becomes an obvious choice for natural control of the prosthetic hand. This requires the classification of sEMG signals to identify the desired finger movements and obtain the command for controlling the prosthetic hand. Some of the earlier attempts to identify finger actions from sEMG were based on an estimate of the amplitude [1] and the rate of change of the sEMG [2]. More recent studies [3-19], have reported significant

developments in the identification of movements for myoelectric control systems.

Researchers have reported success in the use of multiple channels sEMG recording for controlling the prosthetic hand [3-10]. Tenore et al. [7,10] have investigated the effectiveness of different configurations of electrode arrays (19 and 32) on the performance of the prosthetic control, both on able – bodied and trans-radial amputees. However, such systems are complex and the variation in electrode placement during sEMG recording can alter the signal and the outcomes significantly [17,20] making the technology unsuitable for self-administration by the user or their carer. There is also significant variation of sEMG magnitude and spectrum between different experiments due to a number of factors that cannot be controlled [12,21,22]. A single channel system that can reliably identify the finger actions and in which the location of electrodes is not critical, is highly desirable. Smith et al. [8] and Chen et al. [9] attempted to minimise the number of

* Correspondence: sridhar.arjunan@rmit.edu.au

†Equal contributors

¹Bio-signals Lab, School of Electrical and Computer Engineering, RMIT University, GPO Box 2476, Melbourne, Victoria 3001, Australia
Full list of author information is available at the end of the article

electrodes (using six to eight electrodes) to decode four different finger flexions. Another method reported by Englehart et al. [4] is based on dimensionality reduction using principal components analysis after wavelet decomposition. However, overlapping muscles and presence of noise and artefacts makes this a challenging task.

Many researchers have worked to decompose the sEMG signal into constituent MUAPs [23,24]. Pelvin and Zazula [25] reported the use of higher order statistics for decomposing the EMG signal into the fundamental components, the individual motor unit action potentials (MUAP). However, these techniques are based on the priori of shape and density of MUAP, making them unsuitable when there are multiple muscles because the shape of MUAP from different muscles can vary due to the difference in the transmission pathways. SEMG recordings integrate the electrical activity from all adjoining muscles and thus lack muscle selectivity. Further, low-level muscle activity, such as during finger flexion, makes the signal susceptible to noise and artefacts. Different choices of global features of the signal using advanced signal processing and pattern recognition techniques do not address these fundamental issues and such research can at best result in marginal improvement. While the systems reported in literature appear to be suitable for recognising gross movements, they are ineffective for complex movements such as wrist and finger movements [2] where a number of muscles are involved.

This paper reports an alternate method to estimate the relative strength of contraction of different muscles. It is based on the transmission properties of action potentials in body tissues [26-28] and on the assumption that the MUAP are sparse and cause singularities in the recordings. The MUAP arriving from one muscle were considered to have similar magnitudes, and those from other muscles have different magnitudes. Wavelet maxima were used to identify the singularities and cluster analysis was performed to determine the relative strength corresponding to different actions. The densities of the clusters of wavelet maxima were classified using a twin support vector machine (TSVM) to determine the associated finger flexion and the technique has been called Wavelet Maxima Density (WMD) technique.

Methods

Subjects

Eleven healthy subjects (7 male and 4 female; Mean age = 27.4 (± 2.55) years; Mean weight = 68.7 (± 3.56) kg; and Mean height = 169.3 (± 6.2) cm) volunteered to participate in this study. The healthy participants exclusion criterion was; (i) no history of myo or neuropathology, and (ii) no evident abnormal motion restriction. The technique was tested on one amputee participant (Female; Age = 42

Years) who volunteered to take part in this study. Amputee volunteer had trans-radial one-third proximal amputation of the right forearm (with 9 cm long stump).

SEMG recording procedures

Bipolar electrodes (DELSYS, Boston, MA, USA) were placed on the forearm (FDS) muscle for the healthy participants (Figure 1a) in accordance with standard procedures [29] to record surface electromyogram (sEMG). These are active electrodes, with the preamplifier and two electrodes built into a single unit. The electrodes are self-adhesive, and have two silver bars; each of 1 mm thickness, 10 mm length and the fixed inter-electrode distance of 10 mm. Electrolyte Gel (Sigma) were used on the electrodes prior to affixing them on the skin.

Experiments were repeated twice for the able-bodied volunteers, with electrodes placed on the proximal end in the first experiment- EXP1, and on the distal end for the second experiment – EXP2 (Figure 1a). This was done to determine the effect of variation in the electrode location on the outcome of the experiments. The ground electrode




Figure 1 Electrode Placement a) at the distal (EXP1) and proximal end (EXP2) b) on the Trans- radial amputee.

was placed on the volar aspect of the wrist. For the amputee participant, the electrodes were placed on the remaining stump of the participant as shown in Figure 1b.

DELSYS (Boston, MA, USA) sEMG acquisition system was used to record the signal. The system gain was 1000, CMRR was 92 dB, and bandwidth was 20–450 Hz, with 12 dB/ octave roll-off. The input impedance of the system was 115 Pico-farad in parallel with 1 K-ohm. The sampling rate of the system was 1024 samples/ second for each channel and the resolution were 16 bits/ sample. Prior to the placement of electrodes, the skin of the participant was prepared by shaving (if required) and exfoliation to remove dead skin. Skin was cleaned with 70% v/v alcohol swab to remove any oil or dust from the skin surface.

Experimental protocol

Experiments were conducted where the sEMG was recorded while the participants performed four sets of generic finger flexion actions (Figure 2) labelled and described below.

- Background activity: All fingers resting.
- Class 1: Flexion of little (pinkie) finger
- Class 2: Flexion of ring finger
- Class 3: Flexion of middle finger
- Class 4: Flexion of index finger

These generic actions were selected for the following reasons:

- these actions will allow the user to control individual fingers in the recently available advanced

robotic/ prosthetic hand [7,10,19] and utilise these devices to the maximum advantage.

- these actions can be used for communication and control commands that can be used for several different devices and applications.

The participants performed the flexion without any resistance and as was convenient and easily reproducible by them. SEMG was recorded through the experiment. The examiner gave on-screen and oral commands to the participant to perform the action without any fixed order of the fingers. Each flexion was maintained for about 7–8 seconds and was repeated twelve times. The experiments were repeated twice with changed electrode location.

The experimental protocol was approved by the RMIT University Human Ethics Committee and Alfred Health Ethics Committee and performed in accordance with the Declaration of Helsinki 1975, as revised in 2004. Prior to the recording, the participants were encouraged to familiarize themselves with the experimental protocol and with the equipment. For the experiment with the amputee, bilateral action training modality was performed [30]. The amputee subject performed the finger flexions with the healthy hand while attempting to repeat the same flexion with the phantom limb [30].

Data analysis

Removal of noise and background activity

As a first step, the signal corresponding to class 1 for each subject on each day was normalised to the root mean square (RMS) of the recording of the same subject. The




Figure 2 Four finger flexions; (a) Little Finger; (b) Ring Finger; (c) Middle Finger; (d) Index Finger. Proximal inter – phalangeal joint was flexed in these experiments.

next step was the removal of the background activity from the signal (Figure 3). This becomes more challenging because when signal strength is low, the noise magnitude becomes comparable with the signal itself.

The pilot analysis of the recordings showed large inter-experimental variations in the spectrum of the background activity, making stationary spectral filtering unsuitable because of the variations in the noise characteristics [31]. For this reason, adaptive spectral subtraction filters were used to remove the background activity from the recorded signal. The template of the spectrum of the noise was obtained for each subject using a band-pass filter (4th order Butterworth, frequency 10 to 450 Hz) and averaging over 20 windows (>20).

Decomposing sEMG to obtain relative strength of contraction

The first step of the proposed method required the identification of the temporal location of the MUAPs. For this purpose, sEMG signal was decomposed using wavelet transform (bi-orthogonal wavelet 'bior3.3') and the maxima were identified based on the change of sign [32].

Local maxima, $Wf(s, x_n)$ can be described mathematically as follows [32]:

$$Wf(s, x_{n-1}) < Wf(s, x_n) > Wf(s, x_{n+1}) \quad (1)$$

Where $Wf(s, x)$ is the wavelet transform of the function $f(x)$ at a scale s and $n = 2$ to $N-1$, N is the total number of coefficients at any given scale s .

The wavelet maxima that were present in each of the scales and travelled from finest scale to coarsest scale were considered [33,34]. Other wavelet maxima were

rejected as being random transients. The data of each flexion was segmented into 300 ms windows, and the magnitudes at the lowest (finest) scale of the accepted maxima were obtained. Based on the magnitude, the magnitude and number of maxima points of every windowed recording (density of peaks) corresponding to each flexion were the two dimensional representation of the recording. These were separated into four (number of different actions) groups using cluster analysis. The centroid and corresponding density of each cluster was determined for each time segment, and averaged for the duration of each flexion. This resulted in one value of the magnitude and the corresponding density of peaks for each action. This two dimensional feature set was used to train and test the system.

Classification

Twin support vector machines (TSVM) linear kernel classifier [35] was used to classify the features. The advantages of using TSVM is that it solves two related SVMs, one for each class, and generates two separate hyper planes without the assumption that patterns in each class arise from similar distributions. It allows the use of a different kernel for each class which can be separately optimized based on the data. This data dependent kernel optimization for each class is particularly useful in this study. For more details, please refer to [15,35].

The vectors corresponding to the four densities and four magnitudes were the input to the classifier. During the training phase the associated finger flexion was the target. After the system was trained, the system accuracy was validated using ten-fold cross – validation and tested using Type I error (Specificity) and Type II error




Figure 3 Example of the background and the muscle activity.

(Sensitivity) [15]. The training dataset had 100 data points, and testing was done using 30 data points. The training data set and test data set was chosen randomly using random sub-sampling method. Ten-cross validation method was used to determine the accuracy, sensitivity and specificity of the system. Classification was performed on the data from the individual subject. The system was tested individually for the two experiments to determine the inter-experimental variations.

Results

The average classification accuracy for able-bodied subjects is reported in Table 1 and for amputee subject in Table 2. The sensitivity and specificity results for the able bodied subjects are reported in Table 3 and for the amputee subject in Table 4.

Accuracy of identifying sEMG of able bodied participants

From Table 1, the overall accuracy of the detection of flexion of four classes of fingers (digits 2 to 5) using WMD was found to be 93.41(±1.45)% when sEMG was recorded from the distal end (experiment 1) of the FDS muscle. When sEMG was recorded from the proximal end (experiment 2) of the FDS muscle, there was only a small variation (1% decrease) in the overall accuracy (92.4 ± 3.23)%.

Accuracy of identifying sEMG of amputee

The accuracy of identification of movements from the amputee is tabulated in Table 2. The results show that the average accuracy of the detection of flexion of four classes of fingers (digits 2 to 5) performed based on the bilateral learning was found to be 81.87 (± 13.54)% from sEMG electrode location 2 (Figure 1b). The average accuracy was found to be 74.59 (± 12.52)% from the sEMG electrode location 4. The classifier did not converge for recordings from the electrode locations 1 and 3 and hence has not been reported. This demonstrated that the electrode location had a significant effect on the system accuracy for the amputee subject. For further analysis, electrode location 2 was considered.

Table 1 Mean classification accuracy (%) of different classes recorded from distal end (EXP1) and proximal end (EXP2) from able-bodied all subjects using WMD method

Class	Distal end (EXP1)	Proximal end (EXP 2)
	Mean (±SD)%	Mean (±SD)%
Class 1	92.84 (±4.47)	94.23 (±5.08)
Class 2	95.47 (±4.07)	95.96 (±3.08)
Class 3	92.10 (±5.98)	90.37 (±5.58)
Class 4	93.21 (±5.96)	89.06 (±8.17)
Overall mean (±SD)	93.41 (±1.54)	92.4 (±3.23)

Table 2 Mean classification accuracy (%) of different classes recorded from Ch 2 and Ch 4 from amputee subject using WMD method

Class	Channel 2 mean (±SD)%	Channel 4 mean (±SD)%
Class 1	96.67 (±9.44)	71.67 (±8.08)
Class 2	75 (±9.92)	62.42 (±8.74)
Class 3	66.67 (±10.67)	72.15 (±11.24)
Class 4	89.15 (±8.26)	92.14 (±7.52)
Overall mean (±SD)	81.87 (±13.54)	74.59 (±12.52)

Sensitivity and specificity analysis

From Table 3, the sensitivity of the results for able bodied subjects was 94% while the average specificity of all classes was ~ 97% for experiment 1 and ~ 93% for experiment 2. The results indicate low Type I and Type II error. From Table 4, the average specificity was 85% for channel 2 and 84.5% for channel 4, while the average sensitivity was 85% for channel 2 and 83% for channel 4.

Discussion

The results indicate that WMD sEMG signal analysis, which is based on the volume conduction properties of the tissues, can accurately identify individual finger actions. Based on the volume conduction model reported by [26,27], the muscle tissue forms the major portion of the volume conductor. This is an anisotropic conducting medium, having different conductivities along different axes, relatively higher conductivity in the longitudinal direction and slower in the transverse direction. Therefore electric potential travelling in transverse direction of a muscle is attenuated more rapidly as compared to longitudinal direction [26,27]. Thus the action potential originating from co-located muscle fibres would have similar magnitude and the assumption is that the muscles of the forearm have relatively small cross-sectional area. The magnitude of the singularities of the signal is inversely proportional to the distance between the electrodes and the muscle, while the density of the singularities corresponds to the relative number of motor unit action potentials (MUAP), thus corresponding to the strength of muscle contraction.

In this method, sEMG was represented by the wavelet maxima and these were then grouped based on their magnitude into four groups corresponding to the four finger actions. The average magnitude and density of each of the four groups for a time-window of 300 milliseconds [4,13,16] were the input to a support vector machine that was trained to associate these with the finger action. This time window is the permissible delay for real time operation of the system. The method was also

Table 3 Sensitivity (%) and specificity (%) for each class recorded from distal end (EXP1) and proximal end (EXP2) from able-bodied all subjects using WMD method

Class	Distal end (EXP 1)		Proximal end (EXP 2)	
	Sensitivity mean (\pm SD)%	Specificity mean (\pm SD)%	Sensitivity mean (\pm SD)%	Specificity mean (\pm SD)%
Class 1	92 (\pm 0.05)	98 (\pm 0.02)	94 (\pm 0.07)	98 (\pm 0.02)
Class 2	95 (\pm 0.06)	97 (\pm 0.03)	95 (\pm 0.04)	98 (\pm 0.01)
Class 3	92 (\pm 0.06)	98 (\pm 0.03)	91 (\pm 0.05)	96 (\pm 0.02)
Class 4	93 (\pm 0.06)	98 (\pm 0.02)	90 (\pm 0.06)	97 (\pm 0.03)
Overall mean (\pmSD)	93 (\pm0.01)	98 (\pm0.01)	93 (\pm0.02)	97 (\pm0.01)

tested for the 200 ms ($<$ 300 ms) and no significant change in the classification accuracy was observed.

Two sets of experiments with the electrode placement distinctly different between the two experiments were conducted to determine the reproducibility of the technique. The classification accuracy for able-bodied subjects was in the range between 89% and 96%, while for the trans-radial amputee was between 62% and 96%. The lower accuracy for the trans-radial amputee may be due to number of reasons such as disuse the muscles or damage to the muscle [22], or movement of the insertion point, which makes it difficult to identify suitable location for electrode placement.

The recognition accuracy shows that this technique is significantly better compared with the technique reported by Momen et al. ($56 \pm 13\%$) [13], even though the proposed technique has used single channel while Momen et al. [13] have used two channels. The other significance is that this method has identified generic finger actions, which makes this suitable for controlling each finger of a prosthetic hand, while Momen et al. [13] and other researchers [10,11,14,20] have considered user defined actions. The other strength of this technique is that it is repeatable and not sensitive to differences in electrode placement.

The system requires supervised training, while the classification phase is suitable for being automated without manual supervision. The technique could be applied for prosthetic hand control or for human computer interface.

Conclusion

In this study, we have investigated and reported a novel method to estimate the relative strength of contraction of different muscles to identify individual finger flexions. This method is based on the volume transmission properties of action potentials in body tissues [16], leading to the MUAP of different muscles having different amplitudes. During the low level of muscle contraction such as isometric finger flexion, MUAP are sparse and thus the measure of density of MUAP in terms of the amplitude leads to a measure of strength of contraction of different muscles.

Wavelet maxima detection is a suitable technique to identify the local peaks in the sEMG. Identifying the range of the amplitude of peaks using clustering techniques using the wavelet maxima density (WMD) is an indicator of the muscle activity of different muscles. This technique uses WMD as the set of features that can be classified using twin support vector machine (TSVM) to determine the associated finger flexion. The advantage of the TSVM over other classifier techniques is that it is suitable for unbalanced data sets.

The results show that using WMD, single channel EMG is suitable for accurately identifying individual finger flexion. This technique was tested for able bodied and trans-radial amputee subject and also for variations in the placement of the electrodes. This work has demonstrated that it is possible to accurately identify the intent of individual finger flexion using single channel EMG recorded from the surface of the stump of a transradial amputee patient.

Table 4 Sensitivity (%) and specificity (%) for each class recorded from Ch 2 and Ch 4 from amputee subject using WMD method

Class	Channel 2		Channel 4	
	Sensitivity (%)	Specificity (%)	Sensitivity (%)	Specificity (%)
Class 1	90	95	81	80
Class 2	82	81	80	79
Class 3	79	80	84	81
Class 4	89	85	93	92
Overall mean (\pmSD)	85 (\pm0.05)	85 (\pm0.06)	85 (\pm0.05)	83 (\pm0.06)

Competing interests

The authors declare that they have no competing interests.

Authors' contributions

DKK has designed the experiment, discussed and developed the underlying concepts for the technique and finalised the manuscript. SPA has also designed and conducted the experiments on the amputee subject, performed data analysis and written part of the manuscript. VPS has conducted the experiment with able bodies and developed the wavelet based signal processing technique. All authors read and approved the final manuscript.

Acknowledgements

The authors would like to acknowledge Caulfield Alfred Hospital for providing permission to conduct experiment on amputee volunteer and Prof. Jayadeva, for providing us with the Twin SVM classifier. The authors would also like to thank Dr Michael Chou, Mr. David Wilson-Brown and Ms. Karen Roberts of the Amputee Unit, Caulfield Hospital, Alfred Health for providing assistance and help on the conducting experiment with amputee volunteer.

Author details

¹Bio-signals Lab, School of Electrical and Computer Engineering, RMIT University, GPO Box 2476, Melbourne, Victoria 3001, Australia. ²IP Australia, Canberra, Australia.

Received: 30 April 2012 Accepted: 25 May 2013

Published: 7 June 2013

References

1. Dorcas D, Scott RN: **A three state myoelectric control.** *Med Biol Eng* 1966, **4**:367–372.
2. Childress DA: **A myoelectric three state controller using rate sensitivity.** In *Proceedings of 8th International Conference on Medical and Biological Engineering conference(ICMBE)*. Chicago: Palmer House; 1969:4–5.
3. Hudgins B, Parker PA, Scott RN: **A new strategy for multifunction myoelectric control.** *IEEE Trans Biomed Eng* 1993, **40**:82–94.
4. Englehart K, Hudgins B, Parker PA: **A wavelet-based continuous classification scheme for multifunction myoelectric control.** *IEEE Trans Biomed Eng* 2001, **48**(3):302–311.
5. Karlik B, Tokhi O, Musa A: **A fuzzy clustering neural network architecture for multifunction upper-limb prosthesis.** *IEEE Trans Biomed Eng* 2003, **50**(11):1255–1261.
6. Nagata K, Adno K, Magatani K, Yamada M: **A classification method of hand movements using multi channel electrode.** In *Proceedings of 27th annual international conference of the engineering in medicine and biology society*. Shanghai: IEEE; 2005:2375–2378.
7. Tenore F, Ramos A, Fahmy A, Acharya S, Etienne-Cummings R, Thakor NV: **Decoding of individuated finger movements using surface electromyography.** *IEEE Trans Biomed Eng* 2009, **56**(5):1427–1434.
8. Smith JR, Huberdeau D, Tenore F, Thakor NV: **Real-time myoelectric decoding of individual finger movements for a virtual target task.** In *Proceedings of 31st annual IEEE EMBS international conference*. Minneapolis, Minnesota, USA: IEEE; 2009:2376–2379.
9. Chen X, Lantz V, Kong-Qiao W, Zhang-Yan Z, Xu Z, Ji-Hai Y: **Feasibility of building robust surface electromyography-based hand gesture interfaces.** In *Proceedings of 31st annual IEEE EMBS international conference*. Minneapolis, Minnesota, USA: IEEE; 2009:2983–2986.
10. Tenore F, Ramos A, Fahmy A, Acharya S, Etienne-Cummings R, Thakor NV: **Towards the control of individual fingers of a prosthetic hand using surface EMG signals.** In *Proceedings of 29th annual IEEE EMBS international conference*. Lyon, France: IEEE; 2007:6145–6148.
11. Chan FHY, Yong-Sheng Y, Lam FK, Yuan-Ting Z, Parker PA: **Fuzzy EMG classification for prosthesis control.** *IEEE Trans Rehabil Engin* 2000, **8**(3):305–311.
12. Englehart K, Hudgins B, Parker PA, Stevenson M: **Classification of the myoelectric signal using time-frequency based representations.** *Med Engin and Physics* 1999, **21**:431–438.
13. Momen K, Krishnan S, Chau T: **Real time classification of forearm electromyographic signals corresponding to user-selected intentional movements for multifunction prosthesis control.** *IEEE Trans. Neural Systems and Rehab Engin* 2007, **15**(4):535–542.
14. Arjunan SP, Kumar DK: **Decoding subtle forearm flexions using fractal features of surface electromyogram from single and multiple sensors.** *J Neuroeng Rehabil* 2010, **7**:53.
15. Naik GR, Kumar DK, Jayadeva J: **Twin SVM for gesture classification using the surface electromyogram.** *IEEE Trans Inf Technol Biomed* 2010, **14**(2):301–308.
16. Englehart K, Hudgins B: **A robust, real-time control scheme for multifunction myoelectric control.** *IEEE Trans Biomed Eng* 2003, **50**(7):848–854.
17. Kyberd J, Holland OE, Chappel PH, Smith S, Tregdigo R, Bagwell PJ, Snaith M: **Marcus: A two degree of freedom hand prosthesis with hierarchical grip control.** *IEEE Trans Rehabil Eng* 1995, **3**(1):70–76.
18. Cipriani C, Antfolk C, Controzzi M, Lundborg G, Rosen B, Carrozza MC, Sebelius F: **Online myoelectric control of a dexterous hand prosthesis by transradial amputees.** *IEEE Trans Neural Syst Rehabil Eng* 2011, **19**(3):260–270.
19. Cipriani C, Controzzi M, Carrozza MC: **The SmartHand transradial prosthesis.** *J Neuroeng Rehabil* 2011, **8**:29.
20. Cram GS, Kasman J, Holtz J: *Introduction to surface electromyography*. Gaithersburg, Maryland: Aspen Publishers, Inc; 1998.
21. Kumar DK, Alemu M: **Time-frequency analysis of SEMG- with special consideration to interelectrode spacing.** *IEEE Trans Neural Syst Rehabil Eng* 2004, **11**(4):341–345.
22. Kumar DK, Pah ND: **Thresholding wavelet networks for signal classification.** *Int J Wavelets, Multiresolution and Info Process* 2009, **2**:243–261.
23. Zhou P, Rymer WZ, Suresh N, Zhang L: **A study of surface motor unit action potentials in first dorsal interosseus (FDI) muscle.** In *Proceedings of 23rd annual IEEE EMBS international conference*. Istanbul, Turkey: IEEE; 2001:1074–1077.
24. Kleine BU, van Dijk JP, Lapatki BG, Zwarts MJ, Stegeman DF: **Using two-dimensional spatial information in decomposition of surface EMG signals.** *J Electromyogr Kinesiol* 2007, **17**:535–548.
25. Plevin E, Zazula D: **Decomposition of surface EMG signals using non-linear LMS optimisation of higher-order cumulants.** In *Proceedings of the 15th IEEE Symposium on Computer-Based Medical Systems volume 32*. Maribor, Slovenia; 2002:149–154.
26. Blok JH, Stegeman DF, Oosterom AV: **Three-layer volume conductor model and software package for applications in surface electromyography.** *Ann Biomed Eng* 2002, **30**:566–577.
27. Roeleveld K, Blok J, Stegeman DF, Oosterom AV: **Volume conduction models for surface EMG; confrontation with measurements.** *J Electromyogr Kinesiol* 1997, **7**(4):221–232.
28. Merletti R, Botter A, Troiano A, Merlo E, Minetto MA: **Technology and instrumentation for detection and conditioning of the surface electromyographic signal: state of the art.** *Clin Biomech* 2009, **24**:122–134.
29. Hermens HJ, Freriks B, Disselhorst-Klug C, Rau G: **Development of recommendations for SEMG sensors and sensor placement procedures.** *J Electromyogr Kinesiol* 2000, **10**:361–374.
30. Castellini S, Gruppioni E, Davalli A, Sandini G: **Fine detection of grasp force and posture by amputees via surface electromyography.** *J Physiol Paris* 2009, **103**(3–5):255–262.
31. Andrade OA, Nasuto S, Kyberd P, Sweeney- Reed CM, Van Kanijn FR: **EMG signal filtering based on empirical mode decomposition.** *Biomed Signal Proces and Control* 2006, **1**(1):44–55.
32. Mallat S: *A wavelet tour of signal processing*. London: Academic Press; 1999.
33. Abel EW, Meng H, Forster A, Holder D: **Singularity characteristics of needle EMG IP signals.** *IEEE Trans Biomed Eng* 2006, **53**(2):219–225.
34. Arikidis NS, Abel EW, Forster A: **Interscale wavelet maximum—a fine to coarse algorithm for wavelet analysis of the EMG interference pattern.** *IEEE Trans Biomed Eng* 2002, **49**(4):337–344.
35. Jayadeva J, Khemchandani R, Chandra S: **Twin support vector machines for pattern classification.** *IEEE Trans. Pattern Analysis and Machine Intelligence* 2007, **29**(5):905–910.

doi:10.1186/1743-0003-10-50

Cite this article as: Kumar et al.: Towards identification of finger flexions using single channel surface electromyography – able bodied and amputee subjects. *Journal of NeuroEngineering and Rehabilitation* 2013 **10**:50.